

4-Chloroaniline, 2-Naphthylamine Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C(0.1min)->20°C/min->250°C(0.1min)

Column : Rtx-5Sil MS (30m*0.25mm*0.25um)

Carrier gas : He, 1ml/min (Split ratio 20:1)

Injector : Capillary 280°C

Detector : FID 280°C

Injection Volume : 1ul

2. Chromatogram

