



1,3-Propanediol Analysis

1. Analytical Condition

YL 6500 Series

Oven : 70°C(1min)->70°C/min->220°C(10min)

Column : HP-FFAP (30m*0.53mm*1.0um)

Carrier gas : He, 10ml/min (Split ratio 10:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 0.5ul

2. Chromatogram

