



Acetaldehyde Analysis

1. Analytical Condition

YL 6500 Series

Oven : 150°C(1min)->10 °C/min ->180 °C(0.1min)

Column : HP-PLOT Q (30m*0.32mm*20um)

Carrier gas : He 3ml/min (Split ratio 7 : 1)

Injector : Capillary 200 °C

Detector : FID 220°C

Injection Volume : 1ml (Gas)

2. Chromatogram

