

Acetic acid & Butyric acid Analysis

1. Analytical Condition

YL 6500 Series

Oven : 120°C(1min)->15°C/min->200°C(5min)

Column : TR-Wax (30m*0.25mm*0.25um)

Carrier gas : He, 1ml/min (Split ratio 30:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 1ul

2. Chromatogram

