



## Acrylamide Analysis

### 1. Analytical Condition

#### YL 6500 Series

Oven : 100°C(0.1min) -> 20°C/min -> 230°C(5min)

Column : GL Wax (30m\*0.53mm\*1.0um)

Carrier gas : N<sub>2</sub>, 7ml/min (SR 10:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 1ul

### 2. Chromatogram

