

Aromatic Compounds Analysis in Gasoline

1. Analytical Condition

YL 6500 Series

Oven : 60°C (for 2 min)- temp-programed 2°C/min-120 °C(for 5 min)

Column : HP-1(30m x 0.53mmx 2.65um)

TCEP(560mm*0.38mm id)

Carrier gas : He, 10.0 ml/min (Split ratio 11:1)

Injector : Capillary 230°C

Detector : FID 250°C

2. Chromatogram

