

Bacterial Acid Methyl Ester(BAME) Mix

1. Analytical Condition

YL 6500 Series

Oven : 150°C (for 4 min)- temp-programed 4°C/min-250 °C(for 4 min)

Column : TRB-5(30m*0.25mm*0.25um)

Carrier gas : He, 1.0 ml/min (Split ratio 10:1)

Injector : Capillary 270°C

Detector : FID 270°C

Injection Volume : 1.0ul

2. Chromatogram

