

Butylene Diglycol Analysis

1. Analytical Condition

YL 6500 Series GC

Carrier gas : He / 10 mL/min (SR 20:1)

Injector : 280 °C

Oven : 100°C (1 min) -> 20 °C(min) -> 220 °C(1 min)

Detector : FID 280 °C

Column : TC-Wax (30 m × 0.53 mm × 1.00 μm)

Solvent : Methyl Alcohol

2. Chromatogram

