

Butylene glycol Analysis

1. Analytical Condition

YL6500 Series

Oven : 100°C (for 1 min)- temp-programed 10°C/min-200 °C(for 1 min)

Column : HP-5(30m*0.32mm*0.25um)

Carrier gas : He, 1 ml/min (Split ratio 250:1)

Injector : Capillary 260°C

Detector : FID 260°C

Injection Volume : 1.0ul

2. Chromatogram

