

## C10-C16 Hydrocarbon Analysis

### 1. Analytical Condition

#### YL6500 GC

Oven : 100 °C(for 0.1 min) →20 °C/min(temp-programed) →230 °C(for 5 min)

Column : SP-2100 1/8", 10ft

Carrier gas : He, 12 ml/min

Injector : Packed Inj 250 °C

Detector : FID 250 °C

Injection Volume : 1ul Liquid

### 2. Chromatogram

