

Dibutylhydroxytoluene & Dichlorvos(DDVP) Analysis

1. Analytical Condition

YL 6500 Series

Oven : 50°C(for 0.1 min)->10 °C/min->230 °C(for 1 min)

Column : HP-INNOWAX (30m*0.53mm*1.00um)

Carrier gas : He 6ml/min (Split ratio 10 : 1)

Injector : Capillary 230 °C

Detector : FID 250°C

Injection Volume : 1ul

2. Chromatogram

