



Diphenhydramine Hydrochloride analysis

1. Analytical condition

YL 6500 Series

Oven : 200°C (Isothermal)

Column : HP-1 (30m*0.32mm*0.25um)

Carrier gas : He, 4.0 ml/min (Split ratio 6:1)

Injector : Capillary 300°C

Detector : FID 300°C

Dilution with Ethanol, Injection 1ul

2. Chromatogram

