

EPA 604 phenols 11mix Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C (for 5 min)- temp-programed 15 °C/min-280 °C (for 5 min)

Column : IC Cap-1 (30m*0.32mm*1.0um)

Carrier gas : He, 2ml/min (Split ratio 30:1)

Injector : Capillary 270 °C

Detector : FID 270 °C

Injection Volume : 1.0ul (Liquid), 604 Phenols Calibration Mix (cat.# 31029)

2. Chromatogram

