

Monoethanolamine(MEA) Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C(1min)->10°C/min->200°C(1min)

Column : HP-5 (30m*0.32mm*1.0um)

Carrier gas : He, 1ml/min (Split ratio 200:1)

Injector : Capillary 230°C

Detector : FID 250°C

Injection Volume : 0.2ul

2. Chromatogram

