

Gamma-aminopropyl triethoxy silane Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C(2min)->20°C/min->300°C(4min)

Column : HP-5 (30m*0.32mm*1.0um)

Carrier gas : N2, 3ml/min (Split ratio 10:1)

Injector : Capillary 300°C

Detector : FID 300°C

Injection Volume : 1ul

2. Chromatogram

