



## Propylene Glycol, Diethylene Glycol Analysis

### 1. Analytical Condition

#### YL6500 Series

Oven : 40°C (for 1 min)- temp-programed 10°C/min-230 °C(for 5 min)

Column : HP-Innowax(30m\*0.53mm\*1.0um)

Carrier gas : He, 6 ml/min (Split ratio 10:1)

Injector : Capillary 260°C

Detector : FID 260°C

Injection Volume : 1.0ul

### 2. Chromatogram

