

Allethrin Analysis in Insecticide

1. Analytical Condition

YL 6500 Series

Oven : 150°C(0.1min)->10 °C/min ->220 °C(0.1min)

Column : HP-5 (30m*0.32mm*0.25um)

Carrier gas : He 1ml/min (Split ratio 20 : 1)

Injector : Capillary 270 °C

Detector : FID 270°C

Injection Volume : 1ul

2. Chromatogram

