

Phthalthrin Analysis in Insecticide

1. Analytical Condition

YL 6500 Series

Oven : 80°C(for 3 min)->10 °C/min ->280 °C(for 10 min)

Column : HP-5 (30m*0.32mm*1.00um)

Carrier gas : He 3ml/min (Split ratio 10 : 1)

Injector : Capillary 270 °C

Detector : FID 280°C

Injection Volume : 1ul

2. Chromatogram

