

PAHs(Polynuclear Aromatic Hydrocarbons) Analysis

1. Analytical Condition

YL 6500 Series

Oven : 130°C(1min)->4°C/min->290°C(1min)

Column : HP-5 (30m*0.32mm*0.25um)

Carrier gas : He, 3ml/min (Split ratio 10:1)

Injector : Capillary 280°C

Detector : FID 280°C

Injection Volume : Injection volume 2ul (Dilute solvent : ACN)

2. Chromatogram

