

Squalene Analysis

1. Analytical Condition

YL 6500 Series

Oven : 230°C Isothermal

Column : DB-1(30m*0.32mm*0.25um)

Carrier gas : N₂, 3ml/min (SR 20:1)

Injector : Capillary 285°C

Detector : FID 320°C

2. Chromatogram

