

TEOS Purity Analysis (Tetraethyl orthosilicate)

1. Analytical Condition

YL 6500 Series

Oven : 100°C (for 1 min)- temp-programed 15 °C/min-250 °C(for 3 min)

Column : HP-5 (30m*0.32mm*0.25um)

Carrier gas : N₂, 1ml/min (Split ratio 50:1)

Injector : Capillary 250°C

Detector : FID 250°C

Injection Volume : 1.0ul (Liquid)

2. Chromatogram

