



## Thiophene Analysis

### 1. Analytical Condition

#### YL 6500 Series

Oven : 50°C(5min)->10 °C/min-> 100도(3min)

Column : HP-1(30m\*0.32mm\*1.0um)

Carrier gas : He, 3ml/min (SR 30:1)

Injector : Capillary 230°C

Detector : FID 250°C

Sample : Dilute with Toluene, 1ul

### 2. Chromatogram

