

m-, o-, p-Toluidine Analysis

1. Analytical Condition

YL 6500 Series

Oven : 100°C(0.1min)->15°C/min->220°C(1min)

Column : HP-INNOWAX (30m*0.53mm*1.0um)

Carrier gas : He, 6ml/min (Split ratio 30:1)

Injector : Capillary 230°C

Detector : FID 250°C

Injection Volume : Injection 0.5ul (Dilute solvent : ACN)

2. Chromatogram

