

Triethanolamine Analysis

1. Analytical Condition

YL 6500 Series

Oven : 150°C(0.5min)->15°C/min->250°C(3min)

Column : HP-5 (30m*0.32mm*1.0um)

Carrier gas : He, 2ml/min (Split ratio 70:1)

Injector : Capillary 300°C

Detector : FID 300°C

Injection Volume : 0.4ul

2. Chromatogram

