

Water Analysis in Aceton

1. Analytical Condition

YL 6500 Series

Oven : 150°C(isothermal)

Column : Porapak Q

Carrier gas : He, 20ml/min

Injector : Packed 150°C

Detector : TCD 150°C(Sen=6)

Injection Volume : 1ul

2. Chromatogram

