

Water analysis in Aceton

1. Analytical Condition

YL 6500 Series

Oven : 150°C Isothermal

Column : Porapak Q 6ft 80/100mesh SS

Carrier gas : He 20ml/min

Injector : Packed 150°C

Detector : TCD 150°C

Injection Volume : 1ul

2. Chromatogram

