



Ethanol in Hand Soaps

1. Analytical Condition

YL6500 GC

Oven : 50 °C (3 min) -> 15 °C/min -> 150 °C (1 min)

Column : HP – Innowax (30 m * 0.53 mm * 1.00 µm)

Carrier gas : He, 6 mL/min (Split ratio 15 : 1)

Injector : Capillary 230 °C

Detector : FID 250 °C

Injection Volume : 1 µL

2. Chromatogram

