



General Solvents Mixture (MeOH,IPA,MEK,EA)

1. Analytical Condition

YL6500 GC

Oven : 35 °C (5 min) -> 5 °C/min -> 150 °C (5 min)

Column : HP-5 (30 m * 0.32 mm * 1.0 µm)

Carrier gas : He, 1mL/min (Split ratio 50 :1)

Injector : Capillary 230 °C

Detector : FID 250 °C

Headspace : 80 °C, Heating for 15 min , Gas 0.5mL Injection

2. Chromatogram

