

HAL(5-Aminolevulinic acid hexyl ester hydrochloride) Analysis

1. Analytical Condition

YL 9100 HPLC

Column : Atlantis HILIC Silica(4.6*250mm, 5um)

Mobile Phase : MeOH:Water(100mM Ammonium acetate)=50:50

Detector : ELSD (SC53 °C/DT70 °C, F=9)

Flow : 1ml/min

Injection volume : 10 μ l

Oven Temp : 40°C

2. Chromatogram

