

Acetaminophen Analysis

1. Analytical Condition

YL 9100 Series QC Analyzer

Column : SP C18 MG II (4.6*250mm, 5um)

Detector : UV/vis 254nm

Flow rate : 1ml/min

Injection volume : 20 μ l

Oven Temp : 40°C

Mobile phase : Water : MeOH : Acetic acid = 60 : 40 : 0.1 (mix it together in one bottle)

2. Chromatogram

