

Acetone Analysis by using Reversed column

1. Analytical Condition

YL 9100 Series HPLC

Mobile phase : 0.2% Iso-propyl alcohol

Flow-rate : 1 mL/min

Oven : 35°C

Detector : UV 254 nm

Column : C18 (4.6*150 mm, 5 μ m)

Solvent : MeOH

2. Chromatogram

