

Acrylamide Analysis

1. Analytical Condition

YL 9100 HPLC

Column : Atlantis dC18 (4.6*250mm, 5um)

Mobile Phase : ACN 600ml+Water 400ml + H3PO4 1.5ml + TEA 5ml = 1bottle mix

Detector : UV/Vis 210nm

Flow : 0.8ml/min

Injection volume : 10 μ l

Oven Temp : 40°C

2. Chromatogram

