


Acrylic acid Analysis



1. Analytical Condition

YL 9100 HPLC

Column : Waters Spherisorb C8 (4.6*250mm, 5um)

Mobile Phase : ACN : Water = 30:70(0.15% H₃PO₄ included)

Detector : UV/Vis 215nm

Flow : 1.0ml/min

Injection volume : 5 μ l

Oven Temp : 40°C

2. Chromatogram

