



Analysis of Sugars (6 Mix)

1. Analytical Condition

YL9100 HPLC System

Mobile Phase : Acetonitrile : UPW = 75 : 25

Flow rate : 1 mL/min

Column : Cosmosil Sugar - D (4.6 * 250 mm)

Detector : RID, 35 °C

Column Temp. : 35 °C

Injection Volume: 20 µL

2. Chromatogram

