

Benzopyrene Analysis

1. Analytical Condition

YL 9100 Series HPLC

Mobile Phase : ACN:Water=80:20
Flow Rate : 1.0ml/min
Column Oven : 35°C
Detector : FLD Ex 294nm Em 404nm
Column : C18 (4.6*250mm, 5um)
Injection Volumn : 20ul sample loop

2. Chromatogram

