


Bisphenol A Analysis



1. Analytical Condition

YL9100 HPLC

Mobile Phase : ACN : Water = 80 : 20

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : FLD Ex 275/Em 300, UV/Vis 230nm

Column : C18 (4.6*250mm, 5um)

Injection Volume : 20ul

2. Chromatogram

