

# Caffeine Analysis in Black Coffee

## 1. Analytical Condition

### YL 9100 HPLC

Column : Capcell Pak C18 (4.6\*150mm, 5um)


Mobile Phase : ACN : Water = 60 : 40

Detector : UV/Vis 280nm

Flow : 1.0ml/min

Injection volume : 10 $\mu$ l

Oven Temp : 40°C



## 2. Chromatogram

