

# Chlorhexidine Gluconate Solution Analysis

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : ACN : Water = 40 : 60 (include 0.1% Acetic acid )

Flow Rate : 1.3ml/min

Column Oven : 40°C

Detector : UV/Vis 258nm

Column : Atlantis dC18(4.6\*250mm, 5um)

Injection Volume : 10ul

## 2. Chromatogram

