



D-Panthenol Analysis (HILIC Column)

1. Analytical Condition

YL 9100 Series HPLC

Mobile Phase : ACN : Water(0.005M Ammonium Acetate) = 80:20

Flow Rate : 0.5ml/min

Column Oven : 35°C

Detector : UV/Vis Detector-210nm

Column : HILIC (4.6*250mm, 5um)

Injection Volumn : 10ul

2. Chromatogram

