


17 β -Estradiol Analysis



1. Analytical Condition

YL9100 HPLC

Mobile Phase : ACN : Water = 67 : 33 (add 0.1% phosphoric acid)

Flow Rate : 1.0ml/min

Column Oven : 30°C

Detector : FLD Ex210nm/Em315nm

Column : C18 (4.6*150mm, 5 μ m)

Injection Volume : 20 μ l

2. Chromatogram

