

# Glycerine & Acetic anhydride Analysis

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : Water 100%

Flow Rate : 1ml/min

Column Oven : 35°C

Detector : RID

Column : Atlantis dC18 (4.6\*250mm, 5um)

Injection Volume : 20ul

## 2. Chromatogram

