

Hormones Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : ACN : Water = 60 : 40

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : PDA 240nm/280nm

Column : C18 (4.6*250mm, 5um)

Injection Volume : 20ul

2. Chromatogram

