



Mercaptoacetic acid (Thioglycolic acid) Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : 30mM perchloric acid aqueous solution

Flow Rate : 1.0 ml/min

Column Oven : Ambient

Detector : UV/Vis 220nm

Column : C18 Bonus RP (4.6*150mm, 5um)

Injection Volume : 20ul

2. Chromatogram

