

# Oleic acid Analysis in Sunflower Seed

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : MeOH :Water=95:5

Flow Rate : 1.0ml/min

Column Oven : 35°C

Detector : ELSD-70 °C, Range ¼, Gas 1.5L/min

Column : C16NH2 column (4.6\*150mm, 5µm)

Injection Volume : 20µl

## 2. Chromatogram

