

Plating Solution Analysis

1. Analytical Condition

YL 9100 HPLC

Column Atlantis T3 (4.6*150mm, 5um)
 Mobile Phase : A-Water(8ml HClO4 per 1L)
 B-ACN
 Detector : UV/Vis 215nm
 Flow : 1.0ml/min
 Injection volume : 5 μ l
 Oven Temp : 40°C

<Gradient Program>

Time	%A	%B
init	100	0
10min	70	30
16min	70	30
16.1min	100	0

2. Chromatogram

