

# Resveratrol Analysis

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : MeOH : Water = 3 : 1

Flow Rate : 1.0ml/min

Column Oven : 40°C

Detector : UV/Vis 280nm

Column : C18 (4.6\*150mm, 5µm)

Injection Volume : 20µl

## 2. Chromatogram

