

Sorbic acid Analysis

1. Analytical Codnition

YL 9100 Series QC Analyzer

Column : C18 (4.6*250mm, 5um)

Detector : UV/Vis 230nm

Flow rate : 1.0ml/min

Injection volume : 20 μ l

Oven Temp : 40°C

Mobile phase : ACN:Water=40:60(0.1% H₃PO₄)

2. Chromatogram

1. Sorbic acid

