

# Stearyl Glycyrrhetinate Analysis

## 1. Analytical Condition

### YL9100 HPLC

Mobile Phase : A = Water

B = MeOH

Flow Rate : 1ml/min

Column Oven : 40°C

Detector : UV/Vis 254nm

Column : C8 (4.6\*150mm, 5um)

Injection Volume : 10ul

<Gradient Program>

	A	B
0min	30	70
7min	0	100
11min	0	100
11.1min	30	70

## 2. Chromatogram

