

Sucralfate Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : pH=3.1 buffer

(ammonium sulfate 27g / Water 1L, use phosphoric acid to reach up to pH = 3.1)

Flow Rate : 0.4ml/min

Column Oven : 35°C

Detector : RID 35°C

Column : Polarity dC18 (4.6*250mm, 5um)

Injection Volume : 20ul

2. Chromatogram

