



## D-panthenol & Salicylic acid Analysis

### 1. Analytical Condition

#### YL 9100 Series

Mobile phase : 0.5% Triethylamine, 0.3% H<sub>3</sub>PO<sub>4</sub> in water : ACN

=90:10 (premixed)

Flow rate : 1 mL

Column oven : 40°C

Detector : UV 210 nm

Column : Atlantis T3 C18(4.6 \* 150 mm, 5 um)

Injection volume : 20 uL

### 2. Chromatogram

